

choosing 120 Hz, the most comfortable one without loss of performance. However there was a tendency for 90 Hz to be better than 120 Hz in terms of accuracy and ITR that may turn out to make up for somewhat less comfort.

ACKNOWLEDGEMENT

We would like to thank Philip van den Broek for technical assistance, Jos Wittebrood at Mindaffect company for helping in the setup, Tsvetomira Tsoneva for her help and insight and the participants for sharing their time.

REFERENCES

- [1] Gao Shang-kai, Wang Yijun, Gao Xiaorong, Hong Bo. Visual and auditory brain-computer interfaces. *IEEE Transactions on Biomedical Engineering*. 2014;61(5):1436–1447.
- [2] Regan David. *Human brain electrophysiology: evoked potentials and evoked magnetic fields in science and medicine*. 1989.
- [3] Herrmann Christoph S. Human EEG responses to 1–100 Hz flicker: resonance phenomena in visual cortex and their potential correlation to cognitive phenomena. *Experimental Brain Research*. 2001;137(3-4):346–353.
- [4] Diez Pablo F, Mut Vicente A, Perona Enrique M Avila, Leber Eric Laciár. Asynchronous BCI control using high-frequency SSVEP. *Journal of Neuroengineering and Rehabilitation*. 2011;8(1):39.
- [5] Berumen G, Tsoneva T. Steady state visual evoked potentials at the boundaries of visual perception. *Proceedings of the 7th Graz Brain-Computer Interface Conference (2017)*.
- [6] Won Dong-Ok, Hwang Han-Jeong, Dähne Sven, Müller Klaus-Robert, Lee Seong-Whan. Effect of higher frequency on the classification of steady-state visual evoked potentials. *Journal of Neural Engineering*. 2015;13(1):016014.
- [7] Won Dong-Ok, Zhang Hai Hong, Guan Cuntai, Lee Seong-Whan. A BCI speller based on SSVEP using high frequency stimuli design. In: *Systems, Man and Cybernetics (SMC), 2014 IEEE International Conference on*. IEEE. 2014, 1068–1071.
- [8] Gembler Felix, Stawicki Piotr, Rezeika Aya, Saboor Abdul, Benda Mihaly, Volosyak Ivan. Effects of Monitor Refresh Rates on c-VEP BCIs. In: *International Workshop on Symbiotic Interaction*. Springer. 2017, 53–62.
- [9] Başaklar Toygun, İder Yusuf Ziya, Tuncel Yiğit. Effects of High Stimulus Presentation Rate on c-VEP based BCIs. In: *2018 Medical Technologies National Congress (TIPTEKNO)*. IEEE. 2018, 1–4.
- [10] Gold Robert. Optimal binary sequences for spread spectrum multiplexing (Corresp.) *IEEE Transactions on Information Theory*. 1967;13(4):619–621.
- [11] Golomb Solomon W et al. *Shift register sequences*. Aegean Park Press, 1967.
- [12] Oostenveld Robert, Praamstra Peter. The five percent electrode system for high-resolution EEG and ERP measurements. *Clinical Neurophysiology*. 2001;112(4):713–719.
- [13] Thielen Jordy, Broek Philip van den, Farquhar Jason, Desain Peter. Broad-Band visually evoked potentials: re(con)volution in brain-computer interfacing. *PLoS One*. 2015;10(7):e0133797.
- [14] Thielen J, Marsman P, Farquhar J, Desain P. Re(con)volution: Accurate response prediction for broadband evoked potentials-based brain computer interfaces. In: *Brain-Computer Interface Research, 2017*, 35–42.
- [15] Wolpaw Jonathan R, Ramoser Herbert, McFarland Dennis J, Pfurtscheller Gert. EEG-based communication: improved accuracy by response verification. *IEEE Transactions on Rehabilitation Engineering*. 1998;6(3):326–333.
- [16] Schreuder Martijn, Höhne Johannes, Treder Matthias, Blankertz Benjamin, Tangermann Michael. Performance optimization of ERP-based BCIs using dynamic stopping. In: *Engineering in Medicine and Biology Society, EMBC, 2011 Annual International Conference of the IEEE*. IEEE. 2011, 4580–4583.